



North Carolina State Laboratory of Public Health
Environmental Sciences
Inorganic Chemistry
Certificate of Analysis

4312 District Drive
 MSC 1918
 Raleigh, NC 27699-1918
 http://slph.ncpublichealth.com
 Phone: 919-733-7308
 Fax: 919-715-8611

FINAL REPORT

Report to: Environmental Health

Name of System:

HARNETT CO ENVIRONMENTAL HEALTH
 307 CORNELIUS HARNETT BLVD
 LILLINGTON, NC 27546

RONALD ELY
 321 ROBIN HILL RD
 Sanford, NC 27332

EIN: 566000306EH

Delivery: NC Courier

StarLiMS ID: **N260138-0025**

Date Collected: 05/18/2026

Time Collected: 11:56

By: MAGGIE WATKINS

Date Received: 05/19/2026

Time Received: 08:10

Sample Type: New Well

Sampling Point:

Well Permit No.

BRES2301-0003

Sample Source:

Receipt Temp.: 12.0 °C

GPS Number:

Comment: Nitrate/Nitrite/Sulfate sample thermal preservation was initiated, but was not at a temperature of 4.0°C upon receipt by the laboratory.

Profile: New Well I

Analyte	Test Result	Allowable Limit	Unit	Qualifier(s)
Arsenic	<0.001	0.010	mg/L	
Barium	0.148	2.0	mg/L	
Cadmium	<0.0005	0.005	mg/L	
Calcium	9		mg/L	
Chloride	<5	250	mg/L	
Chromium	<0.02	0.10	mg/L	
Copper	<0.01	1.3	mg/L	
Fluoride	0.343	4.00	mg/L	
Iron	0.376	0.300	mg/L	
Lead	<0.003	0.015	mg/L	
Magnesium	4		mg/L	
Manganese	0.029	0.05	mg/L	
Mercury	<0.0004	0.002	mg/L	
Nickel	<0.01	0.1	mg/L	
Nitrate	<1	10.0	mg/L	
Nitrite	<0.1	1.00	mg/L	
pH	8.2		N/A	
Selenium	<0.005	0.05	mg/L	
Silver	<0.01	0.10	mg/L	
Sodium	35.6		mg/L	
Sulfate	10.8	250	mg/L	
Total Alkalinity	101		mg/L	
Total Hardness	37		mg/L	
Zinc	<0.05	5.0	mg/L	

Report Date: 05/29/2026

Reported By:

Marc Komlos



NC DEPARTMENT OF
HEALTH AND HUMAN SERVICES
Division of Public Health

Private Well Information and Use Recommendations

For Inorganic Chemical Contaminants

County:

Sample Name:

Sample ID #:

Reviewer:

TEST RESULTS AND USE RECOMMENDATIONS

1. Your well water meets federal drinking water standards *for inorganic chemicals*. Your water can be used for drinking, cooking, washing, cleaning, bathing, and showering based on the *inorganic chemical results only*. You may have other water sampling results that are not taken into account in this report.

2. The following substance(s) exceeded federal drinking water standards or the North Carolina 2L calculated health levels. The North Carolina Division of Public Health recommends that your well water not be used for drinking and cooking unless you install a water treatment system to remove the checked substance(s). However, it may be used for washing, cleaning, bathing, and showering based on the *inorganic chemical results only*.

- | | | | | | | |
|----------------------------------|------------------------------------|----------------------------------|-----------------------------------|--|-----------------------------------|---------------------------------|
| <input type="checkbox"/> Arsenic | <input type="checkbox"/> Barium | <input type="checkbox"/> Cadmium | <input type="checkbox"/> Chromium | <input type="checkbox"/> Copper | <input type="checkbox"/> Fluoride | <input type="checkbox"/> Iron |
| <input type="checkbox"/> Lead | <input type="checkbox"/> Manganese | <input type="checkbox"/> Mercury | <input type="checkbox"/> Nickel | <input type="checkbox"/> Nitrate/Nitrite | <input type="checkbox"/> Selenium | <input type="checkbox"/> Silver |
| <input type="checkbox"/> Zinc | | | | | | |

3. While your lead levels do not exceed federal or state standards, the North Carolina Division of Public Health has concerns with any detection of lead. Should you have any questions please contact the NC Private Well and Health Program at (919) 707 - 5900.

4. Re-sample for lead and /or copper. Take a first draw and 30-second flush sample inside the house (preferably the kitchen sink) and a first draw and 4 minutes flush sample at the wellhead to determine the source of the lead and/ or copper.

5. The following substance(s) exceeded aesthetic drinking water standards. Your water can be used for drinking, cooking, washing, cleaning, bathing, and showering based on the *inorganic chemical results only*, but aesthetic problems such as bad taste, odor, staining of porcelain, etc. may occur. You may want to install a household water treatment system to address aesthetic problems.

- | | | | | |
|-----------------------------------|---------------------------------|-----------------------------------|--|------------------------------------|
| <input type="checkbox"/> Chloride | <input type="checkbox"/> Copper | <input type="checkbox"/> Fluoride | <input checked="" type="checkbox"/> Iron | <input type="checkbox"/> Manganese |
| <input type="checkbox"/> pH | <input type="checkbox"/> Silver | <input type="checkbox"/> Sulfate | <input type="checkbox"/> Zinc | |

6. a. Sodium levels exceed the U.S. Environmental Protection Agency's (USEPA) Health Advisory level for sodium of 20 mg/l. The North Carolina Division of Public Health recommends that only individuals on no or low sodium-restricted diets not use this water for drinking or cooking. It may be used for washing, cleaning, bathing, and showering based on the *inorganic chemical results only*.

b. Your sodium level exceeds 30 mg/L and may pose aesthetic issues such as bad taste, odor, staining of porcelain, etc.

7. Re-sampling is recommended in _____ months, to reinvestigate _____.